

RAYSTATION 10B

DICOM Conformance Statement ProNova Driver



RayStation

10B

Declaration of conformity



Complies with 93/42/EEC Medical Device Directive as amended by M1 to M5. A copy of the corresponding Declaration of Conformity is available on request.

Copyright

This document contains proprietary information that is protected by copyright. No part of this document may be photocopied, reproduced or translated to another language without prior written consent of RaySearch Laboratories AB (publ).

All Rights Reserved. © 2020, RaySearch Laboratories AB (publ).

Trademarks

RayAdaptive, RayAnalytics, RayBiology, RayCare, RayCloud, RayCommand, RayData, RayIntelligence, RayMachine, RayOptimizer, RayPACS, RayPlan, RaySearch, RaySearch Laboratories, RayStation, RayStore, RayTreat, RayWorld and the RaySearch Laboratories logotype are trademarks of RaySearch Laboratories AB (publ)*.

Third-party trademarks as used herein are the property of their respective owners, which are not affiliated with RaySearch Laboratories AB (publ).

RaySearch Laboratories AB (publ) including its subsidiaries is hereafter referred to as RaySearch.

* Subject to registration in some markets.

1 OVERVIEW

This document specifies the DICOM interface for the treatment management system (TMS) RayTreat ProNova driver with Treatment Delivery Devices (TDD) from ProNova. RayTreat ProNova driver can export data associated to a treatment delivery session such as RT Ion Plans, Beams Delivery Instructions, CT images and RT Structure Sets and receive result for the treatment delivery session such as RT Ion Beams Delivery Results, CT and RT images and Spatial Registration objects.

1.1 NETWORK SERVICES

SOP Class Name	SOP Class UID	Provider of Service (SCP)	User of Service (SCU)
Transfer			
CT Image Storage	1.2.840.10008.5.1.4.1.1.2	Yes	No
RT Ion Beams Treatment Record Storage	1.2.840.10008.5.1.4.1.1.481.9	Yes	No
Spatial Registration (REG) Storage	1.2.840.10008.5.1.4.1.1.66.1	Yes	No
Query/Retrieve			
Study Root Query/Retrieve Information Model – FIND	1.2.840.10008.5.1.4.1.2.2.1	Yes	No
Workflow Management			
Unified Procedure Step - Push SOP Class	1.2.840.10008.5.1.4.34.6.1	Yes	No
Unified Procedure Step - Pull SOP Class	1.2.840.10008.5.1.4.34.6.3	Yes	No
Verification			
Verification SOP Class	1.2.840.10008.1.1	Yes	No

TABLE OF CONTENTS

OVERVIEW	2
NETWORK SERVICES	2
TABLE OF CONTENTS	3
INTRODUCTION	4
REVISION HISTORY	4
AUDIENCE	5
REMARKS	5
TERMS AND DEFINITIONS	5
BASICS OF DICOM COMMUNICATION	6
ABBREVIATIONS	7
REFERENCES	7
NETWORKING	8
IMPLEMENTATION MODEL	8
AE SPECIFICATIONS:	9
MEDIA INTERCHANGE	13
TRANSFORMATION OF DICOM TO CDA	14
SUPPORT OF CHARACTER SETS	15
SECURITY	16
SECURITY PROFILES	16
ASSOCIATION LEVEL SECURITY	16
APPLICATION LEVEL SECURITY	16
ANNEXES	17
IOD CONTENTS	17
DATA DICTIONARY OF PRIVATE ATTRIBUTES	37
CODE TERMINOLOGY AND TEMPLATES	37
GRAYSCALE IMAGE CONSISTENCY	37
STANDARD EXTENDED/SPECIALIZED/PRIVATE SOP CLASSES	37
PRIVATE TRANSFER SYNTAXES	37

3 INTRODUCTION

3.1 REVISION HISTORY

Date	Version	Comment
2018-12-13	1.0	ProNova Driver DCS for RayStation Release 8B
2019-02-14	2.0	ProNova Driver DCS for RayStation Release 8B SP1
2019-05-24	1.0	ProNova Driver DCS for RayStation Release 9A
2019-12-03	1.0	ProNova Driver DCS for RayStation Release 9B
2020-03-04	2.0	ProNova Driver DCS for RayStation Release 9B SP1
2020-05-11	1.0	ProNova Driver DCS for RayStation Release 10A
2020-11-27	1.0	ProNova Driver DCS for RayStation Release 10B

3.1.1 PRONOVA Driver changelog

3.1.1.1 Updates between 2.2.0.0 (RayStation 8B) - 2.3.0.0 (RayStation 8B SP1)

- If an exception is thrown when processing the request queue, the driver application will now be taken down. It should then be automatically restarted by the Windows service manager. Previously, any further requests would instead fail and the driver would have to be restarted manually.
- Validation has been added to ensure that for plans with multiple isocenters, the differences in the delivery positions for each beam are as expected. This is done by comparing to the relative positions of the isocenters for each beam in the plan to the table top displacements in the patient setup sequence and the absolute table top positions in the control point sequence.

3.1.1.2 Updates between 2.3.0.0 (RayStation 8B SP1) - 3.0.0.0 (RayStation 9A)

- More detail in logging of plan validation results.
- Fixed a bug where the recorded `Study Instance UID` of image series acquired during a session would be wrong.
- Reading snout id and snout position from the treatment records received from the device and forwarding them to RayTreat.
- Adding validation to make it impossible to have a second drivers connected to the same database as the first driver if the drivers has somehow been severely misconfigured.
- Blocks check in of sessions if any beam with `Treatment Delivery Type` equal to `TREATMENT` has a `Treatment Machine Name` different from the configured `DeviceName`

3.1.1.3 Updates between 3.0.0.0 (RayStation 9A) - 9.1.0 (RayStation 9B)

- Uses the same version numbering as the corresponding RayStation release.
- Validates the checksum that are produced by RayStation DICOM export when plan is created by a RayStation version above 9.1.0.0 (RayStation 9B) and when the plan has DICOM attribute `Manufacturer` set to `RaySearch Laboratories`.
- When creating delivery plan, a new checksum is recalculated and an additional software version is appended to the DICOM plans `Software Versions` as "RaySearch.Driver.ProNova-9.1.0.0";
- No longer validates that the plan has been asked for when setting session in progress.
- No longer requires machine to exist in MachineDB to be able to create tolerance tables.
- Automatic configuration of driver features, synchronizable through Clinic Settings.
- Logs more session information when first receiving it on the driver.
- Validates the private RaySearch DICOM attribute `Internal Treatment Machine Name` in the RT Ion Beam instead of `Treatment Machine Name` since `Treatment Machine Name` may contain alias not matching machine model name.
- Sends `Plan Label`, `Current Fraction Number` and `Number of Fractions Planned` as part of the `Scheduled Processing Parameters` in the UPS.
- Displays progress percentage in RayTreat while session is in progress.
- UPS and BDI is no longer sent as `CONTINUATION` if no meterset has been delivered, even if a treatment record has been received in a previous session for that fraction.
- Enables parsing of treatment records outside of a session context for offline recording. Validating the `Treatment Machine Name`, `Current Fraction Number`, `Patient ID`, `Patient's Name`, `Patient's Sex`, `Patient's Birth Date` and the `Referenced SOP Instance UID` of the `Referenced RT Plan`, in the Treatment Record.
- Now prefers the `Study Instance UID` of the Treatment Record to be the same as the plan that was delivered, also prefers the `Referenced Series Sequence` of the treatment record to contain a reference to the plan that was delivered. If set, this will enable better possibilities for offline treatment recording.
- Improved readability of DCS. Shows the correct indentation of attributes inside a sequence. Removes all attributes from the "Created SOP Instance(s)" chapter where the value is just read but never written. Type 1 values that are not actually read by us now has the comment "Value not read".

3.1.1.4 Updates between 9.1.0 (RayStation 9B) - 9.2.0 (RayStation 9B SP1)

- UPS for a session with a previous treatment record of 0 delivered MU now sends its treatment records in the `Input Information Sequence` as part of the UPS.
- Table top positions are now included when delivering QA session.

3.1.1.5 Updates between 9.2.0 (RayStation 9B SP1) - 10.0.0 (RayStation 10A)

- No changes affecting driver in this release.

3.1.1.6 Updates between 10.0.0 (RayStation 10A) - 10.1.0 (RayStation 10B)

- Setting `MoveOriginatorMessageID` and `MoveOriginatorApplicationEntityTitle` on all CSTORE requests which originates from a CMOVE.
- Parses out the `CurrentFractionNumber` for all beams in the treatment records and always expects all to be the same.
- Introduced `IsClinical` concept to the driver. A Clinical driver can never communicate with non-clinical `RayTreat`. Or vice versa.
- Made connection between driver and `RayTreat` more secure by forcing the usage of HTTPS for a clinical driver.
- New RaaS service now handles the data synchronization with the driver (all drivers will be connected to the same RaaS service). Driver will also always send the PACS data to this service instead of to `RayTreat`.
- Fixed bug related to when multiple CMOVEs was processed simultaneously.

3.2 AUDIENCE

This document is written for users that need to understand how ProNova will integrate into their healthcare facility. This includes both those responsible for overall imaging network policy and architecture, as well as integrators who need to have a detailed understanding of the DICOM features of the product. This document contains some basic DICOM definitions so that any reader may understand how this product implements DICOM features. However, integrators are expected to fully understand all the DICOM terminology, how the tables in this document relate to the product's functionality, and how that functionality integrates with other devices that support compatible DICOM features.

3.3 REMARKS

This document is written for users that need to understand how RayTreat ProNova Driver will integrate into their healthcare facility. This includes both those responsible for overall imaging network policy and architecture, as well as integrators who need to have a detailed understanding of the DICOM features of the product. This document contains some basic DICOM definitions so that any reader may understand how this product implements DICOM features. However, integrators are expected to fully understand all the DICOM terminology, how the tables in this document relate to the product's functionality, and how that functionality integrates with other devices that support compatible DICOM features.

This Conformance Statement is not supposed to replace validation with other DICOM equipment to ensure proper exchange of intended information. In fact, the user should be aware of the following important issues:

- The comparison of different Conformance Statements is just the first step towards assessing interconnectivity and interoperability between the product and other DICOM conformant equipment.
- Test procedures should be defined and executed to validate the required level of interoperability with specific compatible DICOM equipment, as established by the healthcare facility.

3.3.1 Interoperability validation needed

When using RayTreat ProNova Driver together with other software, the DICOM conformance statements must be compared and relevant validation tests run. The DICOM standard by itself does not guarantee interoperability. The Conformance Statement does, however, facilitate a first-level comparison for interoperability between different applications supporting compatible DICOM functionality. RaySearch is also active within the IHE-RO. Contact RaySearch for more info regarding adherence to IHE-RO profiles.

3.3.2 DICOM revision

The module tables listed in the last two chapters are based on part 3 of the DICOM-standard edition 2020a. For extra clarity all attributes in the referenced modules have been listed, even the ones that are not used by ProNova.

3.4 TERMS AND DEFINITIONS

Informal definitions are provided for the following terms used in this Conformance Statement. The DICOM Standard is the authoritative source for formal definitions of these terms.

Abstract Syntax – the information agreed to be exchanged between applications, generally equivalent to a Service/Object Pair (SOP) Class. Examples: Verification SOP Class, Modality Worklist Information Model Find SOP Class, Computed Radiography Image Storage SOP Class.

Application Entity (AE) – an end point of a DICOM information exchange, including the DICOM network or media interface software; i.e., the software that sends or receives DICOM information objects or messages. A single device may have multiple Application Entities.

Application Entity Title – the externally known name of an Application Entity, used to identify a DICOM application to other DICOM applications on the network.

Application Context – the specification of the type of communication used between Application Entities. Example: DICOM network protocol.

Association – a network communication channel set up between Application Entities.

Attribute – a unit of information in an object definition; a data element identified by a tag. The information may be a complex data structure (Sequence), itself composed of lower level data elements. Examples: Patient ID (0010,0020), Accession Number (0008,0050), Photometric Interpretation (0028,0004), Procedure Code Sequence (0008,1032).

Information Object Definition (IOD) – the specified set of Attributes that comprise a type of data object; does not represent a specific instance of the data object, but rather a class of similar data objects that have the same properties. The Attributes may be specified as Mandatory (Type 1), Required but possibly unknown (Type 2), or Optional (Type 3), and there may be conditions associated with the use of an Attribute (Types 1C and 2C). Examples: MR Image IOD, CT Image IOD, Print Job IOD.

Joint Photographic Experts Group (JPEG) – a set of standardized image compression techniques, available for use by DICOM applications.

Module – a set of Attributes within an Information Object Definition that are logically related to each other. Example: Patient Module includes Patient Name, Patient ID, Patient Birth Date, and Patient Sex.

Negotiation – first phase of Association establishment that allows Application Entities to agree on the types of data to be exchanged and how that data will be encoded.

Protocol Data Unit (PDU) – a packet (piece) of a DICOM message sent across the network. Devices must specify the maximum size packet they can receive for DICOM messages.

Service Class Provider (SCP) – role of an Application Entity that provides a DICOM network service; typically, a server that performs operations requested by another Application Entity (Service Class User). Examples: Picture Archiving and Communication System (image storage SCP, and image query/retrieve SCP), Radiology

Information System (modality worklist SCP).

Service Class User (SCU) – role of an Application Entity that uses a DICOM network service; typically, a client. Examples: imaging modality (image storage SCU, and modality worklist SCU), imaging workstation (image query/retrieve SCU)

Service/Object Pair (SOP) Class – the specification of the network or media transfer (service) of a particular type of data (object); the fundamental unit of DICOM interoperability specification. Examples: Ultrasound Image Storage Service, Basic Grayscale Print Management.

Service/Object Pair (SOP) Instance – an information object; a specific occurrence of information exchanged in a SOP Class. Examples: a specific x-ray image.

Tag – a 32-bit identifier for a data element, represented as a pair of four digit hexadecimal numbers, the “group” and the “element”. If the “group” number is odd, the tag is for a private (manufacturer-specific) data element. Examples: (0010,0020) [Patient ID], (07FE,0010) [Pixel Data], (0019,0210) [private data element]

Transfer Syntax – the encoding used for exchange of DICOM information objects and messages. Examples: JPEG compressed (images), little endian explicit value representation.

Unique Identifier (UID) – a globally unique “dotted decimal” string that identifies a specific object or a class of objects; an ISO-8824 Object Identifier. Examples: Study Instance UID, SOP Class UID, SOP Instance UID.

Value Representation (VR) – the format type of an individual DICOM data element, such as text, an integer, a person’s name, or a code. DICOM information objects can be transmitted with either explicit identification of the type of each data element (Explicit VR), or without explicit identification (Implicit VR); with Implicit VR, the receiving application must use a DICOM data dictionary to look up the format of each data element.

3.5 BASICS OF DICOM COMMUNICATION

This section describes terminology used in this Conformance Statement for the non-specialist. This section is not a substitute for training about DICOM, and it makes many simplifications about the meanings of DICOM terms.

Two Application Entities (devices) that want to communicate with each other over a network using DICOM protocol must first agree on several things during an initial network “handshake”. One of the two devices must initiate an Association (a connection to the other device), and ask if specific services, information, and encoding can be supported by the other device (Negotiation).

DICOM specifies a number of network services and types of information objects, each of which is called an Abstract Syntax for the Negotiation. DICOM also specifies a variety of methods for encoding data, denoted Transfer Syntaxes. The Negotiation allows the initiating Application Entity to propose combinations of Abstract Syntax and Transfer Syntax to be used on the Association; these combinations are called Presentation Contexts. The receiving Application Entity accepts the Presentation Contexts it supports.

For each Presentation Context, the Association Negotiation also allows the devices to agree on Roles – which one is the Service Class User (SCU - client) and which is the Service Class Provider (SCP - server). Normally the device initiating the connection is the SCU, i.e., the client system calls the server, but not always.

The Association Negotiation finally enables exchange of maximum network packet (PDU) size, security information, and network service options (called Extended Negotiation information). The Application Entities, having negotiated the Association parameters, may now commence exchanging data. Common data exchanges include queries for worklists and lists of stored images, transfer of image objects and analyses (structured reports), and sending images to film printers. Each exchangeable unit of data is formatted by the sender in accordance with the appropriate Information Object Definition, and sent using the negotiated Transfer Syntax. There is a Default Transfer Syntax that all systems must accept, but it may not be the most efficient for some use cases. Each transfer is explicitly acknowledged by the receiver with a Response Status indicating success, failure, or that query or retrieve operations are still in process.

Two Application Entities may also communicate with each other by exchanging media (such as a CD-R). Since there is no Association Negotiation possible, they both use a Media Application Profile that specifies “pre-negotiated” exchange media format, Abstract Syntax, and Transfer Syntax.

3.6 ABBREVIATIONS

Name	Meaning
AE	Application Entity
CT	Computed Tomography
DICOM	Digital Imaging and Communications in Medicine
IHE / IHE-RO	Integrating the Healthcare Enterprise. IHE-RO deals with integrating Radiation Oncology.
IOD	Information Object Definition
JPEG	Joint Photographic Experts Group
MR	Magnetic Resonance Imaging
PACS	Picture Archiving and Communication System
PET	Positron Emission Tomography
PTS	Proton Planning System (used by IBA)
RT	Radiotherapy
SCP	Service Class Provider
SCU	Service Class User
SOP	Service-Object Pair
TDD	Treatment Delivery Device
TMS	Treatment Management System
TPS	Treatment Planning System

3.7 REFERENCES

- NEMA PS3 Digital Imaging and Communications in Medicine (DICOM) Standard, available free at <http://medical.nema.org/>


4 NETWORKING

The following diagram illustrates the application dataflow between RayTreat ProNova Driver and the specific TDD.

4.1 IMPLEMENTATION MODEL

4.1.1 Application data flow

The following diagram illustrates the application data flow between RayTreat ProNova Driver and the specific TDD.



The scenario starts with a C-FIND query for Unified Procedure Steps from the remote client. The client can then take responsibility for the UPS by setting it to IN PROGRESS. Once the UPS is IN PROGRESS the RayTreat ProNova Driver only allows requests corresponding to the current session until the session has been completed in the application.

4.1.2 Functional Definition of AEs

4.1.2.1 Functional Definition of "ProNova Application Entity"

The following operations are supported:

CT Image

- C-STORE for setup CT images.
- C-MOVE for planning CT images.

Spatial Registration (SR0)

- C-STORE for registration between setup and planning images.

RT Image

- C-STORE for setup RT Images.

RT Structure Set

- C-STORE for RT Structure Set for the isocenter of the MVCT volume.

Beams Delivery Instructions

- C-MOVE for BDIs related to the Unified Procedure Step.

RT Ion Plan

- C-MOVE for RT Ion Plan related to the Unified Procedure Step.

RT Ion Beams Treatment Record

- C-STORE for delivery result.

Modality Performed Procedure Step - PULL

- C-FIND for worklist query.
- N-ACTION for UPS status changes.
- N-SET for progress update.

Verification

- C-ECHO for connection verification

4.1.3 Sequence of Real World Activities

4.1.3.1 Prepare session

Once the patient is checked in to the session, Unified Procedure Steps will be created and available for Worklist queries.

4.1.3.2 Manual cancellation

The procedure step can be canceled by the user in the application. Further requests relation to the session will be rejected.

4.1.3.3 Complete session

All sessions, including canceled sessions, needs to be completed by the user in the application before another session can be started.

4.2 AE SPECIFICATIONS:

4.2.1 RayTreat ProNova Driver Application Entity

4.2.1.1 SOP Classes

SOP Class Name	SOP Class UID	Provider of Service (SCP)	User of Service (SCU)
Transfer			
CT Image Storage	1.2.840.10008.5.1.4.1.1.2	Yes	No
RT Ion Beams Treatment Record Storage	1.2.840.10008.5.1.4.1.1.481.9	Yes	No
Spatial Registration (REG) Storage	1.2.840.10008.5.1.4.1.1.66.1	Yes	No
Query/Retrieve			
Study Root Query/Retrieve Information Model – FIND	1.2.840.10008.5.1.4.1.2.2.1	Yes	No
Workflow Management			
Unified Procedure Step - Push SOP Class	1.2.840.10008.5.1.4.34.6.1	Yes	No
Unified Procedure Step - Pull SOP Class	1.2.840.10008.5.1.4.34.6.3	Yes	No
Verification			
Verification SOP Class	1.2.840.10008.1.1	Yes	No

4.2.1.2 Association Policies

Not applicable

4.2.1.3 General

The DICOM standard Application context shall be specified.

Application Context Name	1.2.840.10008.3.1.1.1
--------------------------	-----------------------

The maximum PDU size in not configurable and is set to 16384 for SCU and unlimited for SCP.

4.2.1.4 Number of Associations

Any number of incoming concurrent associations are accepted.

4.2.1.5 Asynchronous Nature

RayTreat ProNova Driver does not support asynchronous communication (multiple outstanding transactions over a single Association).

4.2.1.6 Implementation Identity Information

Not applicable

4.2.1.7 Association Initiation Policy

The implementation for this Application Entity is:

Implementation Class UID	1.3.6.1.4.1.30071.8
Implementation Version Name	fo-dicom-raysearch 4.0.4 (based on official fo-dicom 4.0.2)

4.2.1.8 Activity C-ECHO

4.2.1.8.1 Description and Sequencing of Activities

A C-ECHO request can always be sent to the ProNova driver.

4.2.1.8.2 Accepted Presentation Context

Abstract Syntax Name	Abstract Syntax UID	Transfer Syntax name	Transfer Syntax UID	Role	Extended Negotiation
Verification	1.2.840.10008.1.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1		
		Explicit VR Big Endian	1.2.840.10008.1.2.2		

4.2.1.9 Activity C-FIND

4.2.1.9.1 Description and Sequencing of Activities

Not applicable

4.2.1.9.2 Accepted Presentation Context

Abstract Syntax Name	Abstract Syntax UID	Transfer Syntax name	Transfer Syntax UID	Role	Extended Negotiation
Unified Procedure Step - Pull SOP Class	1.2.840.10008.5.1.4.34.6.3	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1		
		Explicit VR Big Endian	1.2.840.10008.1.2.2		

4.2.1.10 Activity C-MOVE

4.2.1.10.1 Description and Sequencing of Activities

Not applicable

4.2.1.10.2 Accepted Presentation Context

Abstract Syntax Name	Abstract Syntax UID	Transfer Syntax name	Transfer Syntax UID	Role	Extended Negotiation
Study Root Query/Retrieve Information Model - MOVE	1.2.840.10008.5.1.4.1.2.2.2	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1		
		Explicit VR Big Endian	1.2.840.10008.1.2.2		

4.2.1.11 Activity C-STORE

4.2.1.11.1 Description and Sequencing of Activities

Not applicable

4.2.1.11.2 Accepted Presentation Context

Abstract Syntax Name	Abstract Syntax UID	Transfer Syntax name	Transfer Syntax UID	Role	Extended Negotiation
CT Image Storage	1.2.840.10008.5.1.4.1.1.2	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1		
		Explicit VR Big Endian	1.2.840.10008.1.2.2		
RT Image Storage	1.2.840.10008.5.1.4.1.1.481.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None

		Explicit VR Little Endian	1.2.840.10008.1.2.1		
		Explicit VR Big Endian	1.2.840.10008.1.2.2		
RT Structure Set Storage	1.2.840.10008.5.1.4.1.1.481.3	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1		
		Explicit VR Big Endian	1.2.840.10008.1.2.2		
Spatial Registration Storage	1.2.840.10008.5.1.4.1.1.66.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1		
		Explicit VR Big Endian	1.2.840.10008.1.2.2		
RT Ion Beams Treatment Record Storage	1.2.840.10008.5.1.4.1.1.481.9	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1		
		Explicit VR Big Endian	1.2.840.10008.1.2.2		

4.2.1.11.3 Status Response

Service Status	Further meaning	Error Code	Reason
Failure	Storage Cannot Understand	Cxxx	Cannot find session or validation failed.
	SOP class not supported	0122	SOP class not supported.
Success	Success	0000	

4.2.1.12 Activity N-ACTION

4.2.1.12.1 Description and Sequencing of Activities

Not applicable

4.2.1.12.2 Accepted Presentation Context

Abstract Syntax Name	Abstract Syntax UID	Transfer Syntax name	Transfer Syntax UID	Role	Extended Negotiation
Unified Procedure Step - Pull SOP Class	1.2.840.10008.5.1.4.34.6.3	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1		
		Explicit VR Big Endian	1.2.840.10008.1.2.2		

4.2.1.12.3 Status Response

Service Status	Further meaning	Error Code	Reason
Refused	QueryRetrieveUnableToProcess	C000	Unknown SOP Instance UID.
	NoLongerUpdateUps	C300	The UPS may no longer be updated.
	IncorrectTransactionUid	C301	The correct Transaction UID was not provided.
	AlreadyInProgress	C302	The UPS is already IN PROGRESS
	IncorrectTransactionUid	C301	The correct Transaction UID was not provided.
	SopInstanceUidDoesNotExists	C307	Specified SOP Instance UID does not exist or is nota UPS Instance managed by this SCP
Failure	AlreadyCanceled	B304	The UPS is already in the requested state of CANCELED.
	AlreadyCompleted	B306	The UPS is already in the requested state of COMPLETED.
Success	Success	0000	

4.2.1.13 Activity N-SET

4.2.1.13.1 Description and Sequencing of Activities

Not applicable

4.2.1.13.2 Accepted Presentation Context

Abstract Syntax Name	Abstract Syntax UID	Transfer Syntax name	Transfer Syntax UID	Role	Extended Negotiation
Unified Procedure Step - Pull SOP Class	1.2.840.10008.5.1.4.34.6.3	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1		

		Explicit VR Big Endian	1.2.840.10008.1.2.2		
--	--	------------------------	---------------------	--	--

4.2.1.13.3 Status Response

Service Status	Further meaning	Error Code	Reason
Refused	QueryRetrieveUnableToProcess	C000	Unknown SOP Instance UID.
	NoLongerUpdateUps	C300	The UPS may no longer be updated.
	IncorrectTransactionUid	C301	The correct Transaction UID was not provided.
	SopInstanceUidDoesNotExists	C307	Specified SOP Instance UID does not exist or is nota UPS Instance managed by this SCP
Success	Success	0000	

5 MEDIA INTERCHANGE

Not applicable

6 TRANSFORMATION OF DICOM TO CDA

Not applicable

7 SUPPORT OF CHARACTER SETS

RayTreat ProNova Driver support the following charactersets in addition to the default

- ISO_IR 192

8 SECURITY

8.1 SECURITY PROFILES

No Security Profiles are supported.

8.2 ASSOCIATION LEVEL SECURITY

RayTreat ProNova Driver checks the following values for validation of received Association Open Requests:

- Called AE Title.

8.3 APPLICATION LEVEL SECURITY

None supported.

9 ANNEXES

9.1 IOD CONTENTS

9.1.1 Created SOP Instance(s)

9.1.1.1 RT Ion Plan IOD

IE	Module	Used
Patient	Patient Module	No
Study	General Study Module	Yes
Series	RT Series Module	Yes
Frame of Reference	Frame of Reference Module	No
Equipment	General Equipment Module	Yes
Plan	RT General Plan Module	Yes
	RT Ion Tolerance Tables Module	Yes
	RT Ion Beams Module	Yes
	SOP Common Module	Yes

9.1.1.1.1 General Study Module

Attribute name	Tag	Vr	Type	Comment
Study ID	(0020,0010)	SH	2	

9.1.1.1.2 RT Series Module

Attribute name	Tag	Vr	Type	Comment
Series Instance UID	(0020,000E)	UI	1	

9.1.1.1.3 General Equipment Module

Attribute name	Tag	Vr	Type	Comment
Software Versions	(0018,1020)	LO	3	When delivery plan is created, version of the driver is appended.

9.1.1.1.4 RT General Plan Module

Attribute name	Tag	Vr	Type	Comment
Plan Intent	(300A,000A)	CS	3	Always VERIFICATION.
Referenced RT Plan Sequence	(300C,0002)	SQ	3	
>Referenced SOP Class UID	(0008,1150)	UI	1	
>Referenced SOP Instance UID	(0008,1155)	UI	1	
>RT Plan Relationship	(300A,0055)	CS	1	Always PREDECESSOR.

9.1.1.1.5 RT Ion Tolerance Tables Module

Attribute name	Tag	Vr	Type	Comment
Ion Tolerance Table Sequence	(300A,03A0)	SQ	1	
>Tolerance Table Number	(300A,0042)	IS	1	
>Tolerance Table Label	(300A,0043)	SH	3	
>Gantry Angle Tolerance	(300A,0044)	DS	3	
>Beam Limiting Device Angle Tolerance	(300A,0046)	DS	3	
>Beam Limiting Device Tolerance Sequence	(300A,0048)	SQ	3	
>>RT Beam Limiting Device Type	(300A,00B8)	CS	1	
>>Beam Limiting Device Position Tolerance	(300A,004A)	DS	1	

>Patient Support Angle Tolerance	(300A,004C)	DS	3	
>Table Top Vertical Position Tolerance	(300A,0051)	DS	3	
>Table Top Longitudinal Position Tolerance	(300A,0052)	DS	3	
>Table Top Lateral Position Tolerance	(300A,0053)	DS	3	
>Table Top Pitch Angle Tolerance	(300A,004F)	FL	3	
>Table Top Roll Angle Tolerance	(300A,0050)	FL	3	

9.1.1.1.6 RT Ion Beams Module

Attribute name	Tag	Vr	Type	Comment
Ion Beam Sequence	(300A,03A2)	SQ	1	
>Referenced Tolerance Table Number	(300C,00A0)	IS	3	
>Planned Verification Image Sequence	(300A,00CA)	SQ	3	
>>Imaging Device-Specific Acquisition Parameters	(300A,00CC)	LO	3	

9.1.1.1.7 SOP Common Module

Attribute name	Tag	Vr	Type	Comment
SOP Instance UID	(0008,0018)	UI	1	
RaySearch Private Creator	(4001,0010)	LO	3	RAYSEARCHLABS 2.0
RaySearch Checksum Algorithm Version	(4001,1060)	LO	3	Set on delivery plan.
RaySearch Checksum Data	(4001,1061)	OB	3	Set on delivery plan.

9.1.1.2 RT Ion Beams Treatment Record IOD

IE	Module	Used
Patient	Patient Module	Yes
Study	General Study Module	Yes
Series	RT Series Module	Yes
Equipment	General Equipment Module	Yes
Treatment Record	RT General Treatment Record Module	Yes
	RT Treatment Machine Record Module	Yes
	RT Ion Beams Session Record Module	Yes
	SOP Common Module	Yes

9.1.1.2.1 Patient Module

Attribute name	Tag	Vr	Type	Comment
Patient's Name	(0010,0010)	PN	2	
Patient ID	(0010,0020)	LO	2	
Patient's Birth Date	(0010,0030)	DA	2	
Patient's Sex	(0010,0040)	CS	2	

9.1.1.2.2 General Study Module

Attribute name	Tag	Vr	Type	Comment
Study Instance UID	(0020,000D)	UI	1	
Study Date	(0008,0020)	DA	2	
Study Time	(0008,0030)	TM	2	
Referring Physician's Name	(0008,0090)	PN	2	
Study ID	(0020,0010)	SH	2	

Study Description	(0008,1030)	LO	3	
-------------------	-------------	----	---	--

9.1.1.2.3 RT Series Module

Attribute name	Tag	Vr	Type	Comment
Modality	(0008,0060)	CS	1	Always RTRECORD.
Series Instance UID	(0020,000E)	UI	1	
Series Number	(0020,0011)	IS	2	Set to 1

9.1.1.2.4 General Equipment Module

Attribute name	Tag	Vr	Type	Comment
Manufacturer	(0008,0070)	LO	2	Set to RaySearch Laboratories
Manufacturer's Model Name	(0008,1090)	LO	3	Set to RayTreatment
Software Versions	(0018,1020)	LO	3	

9.1.1.2.5 RT General Treatment Record Module

Attribute name	Tag	Vr	Type	Comment
Instance Number	(0020,0013)	IS	1	Set to 1
Treatment Date	(3008,0250)	DA	2	
Treatment Time	(3008,0251)	TM	2	
Referenced RT Plan Sequence	(300C,0002)	SQ	2	
>Referenced SOP Class UID	(0008,1150)	UI	1	
>Referenced SOP Instance UID	(0008,1155)	UI	1	
Referenced Treatment Record Sequence	(3008,0030)	SQ	3	
>Referenced SOP Class UID	(0008,1150)	UI	1	
>Referenced SOP Instance UID	(0008,1155)	UI	1	

9.1.1.2.6 RT Treatment Machine Record Module

Attribute name	Tag	Vr	Type	Comment
Treatment Machine Sequence	(300A,0206)	SQ	1	
>Manufacturer	(0008,0070)	LO	2	
>Manufacturer's Model Name	(0008,1090)	LO	2	
>Device Serial Number	(0018,1000)	LO	2	

9.1.1.2.7 RT Ion Beams Session Record Module

Attribute name	Tag	Vr	Type	Comment
Number of Fractions Planned	(300A,0078)	IS	2	
Primary Dosimeter Unit	(300A,00B3)	CS	1	Always MU.
Treatment Session Ion Beam Sequence	(3008,0021)	SQ	1	
>Referenced Beam Number	(300C,0006)	IS	1	
>Beam Name	(300A,00C2)	LO	1	
>Beam Type	(300A,00C4)	CS	1	
>Radiation Type	(300A,00C6)	CS	1	
>Scan Mode	(300A,0308)	CS	1	
>Number of Wedges	(300A,00D0)	IS	1	
>Number of Compensators	(300A,00E0)	IS	1	
>Recorded Compensator Sequence	(3008,00C0)	SQ	1C	
>>Referenced Compensator Number	(300C,00D0)	IS	1	

>>Compensator ID	(300A,00E5)	SH	3	
>Number of Boli	(300A,00ED)	IS	1	
>Referenced Bolus Sequence	(300C,00B0)	SQ	1C	
>>Referenced ROI Number	(3006,0084)	IS	1	
>Number of Blocks	(300A,00F0)	IS	1	
>Recorded Block Sequence	(3008,00D0)	SQ	1C	
>>Referenced Block Number	(300C,00E0)	IS	1	
>Recorded Snout Sequence	(3008,00F0)	SQ	1C	
>>Snout ID	(300A,030F)	SH	1	
>Applicator Sequence	(300A,0107)	SQ	1C	
>>Applicator ID	(300A,0108)	SH	1	
>>Applicator Type	(300A,0109)	CS	1	
>Number of Range Shifters	(300A,0312)	IS	1	
>Recorded Range Shifter Sequence	(3008,00F2)	SQ	1C	
>>Referenced Range Shifter Number	(300C,0100)	IS	1	
>>Range Shifter ID	(300A,0318)	SH	1	
>Number of Lateral Spreading Devices	(300A,0330)	IS	1	
>Recorded Lateral Spreading Device Sequence	(3008,00F4)	SQ	1C	
>>Referenced Lateral Spreading Device Number	(300C,0102)	IS	1	
>>Lateral Spreading Device ID	(300A,0336)	SH	1	
>Number of Range Modulators	(300A,0340)	IS	1	
>Recorded Range Modulator Sequence	(3008,00F6)	SQ	1C	
>>Referenced Range Modulator Number	(300C,0104)	IS	1	
>>Range Modulator Type	(300A,0348)	CS	1	
>>Beam Current Modulation ID	(300A,034C)	SH	1C	
>Patient Support Type	(300A,0350)	CS	1	
>Current Fraction Number	(3008,0022)	IS	2	
>Treatment Delivery Type	(300A,00CE)	CS	2	
>Treatment Termination Status	(3008,002A)	CS	1	Always UNKNOWN.
>Treatment Verification Status	(3008,002C)	CS	2	
>Specified Primary Meterset	(3008,0032)	DS	3	
>Delivered Primary Meterset	(3008,0036)	DS	3	
>Number of Control Points	(300A,0110)	IS	1	Set to 2
>Ion Control Point Delivery Sequence	(3008,0041)	SQ	1	
>>Referenced Control Point Index	(300C,00F0)	IS	1	
>>Treatment Control Point Date	(3008,0024)	DA	1	
>>Treatment Control Point Time	(3008,0025)	TM	1	
>>Specified Meterset	(3008,0042)	DS	2	
>>Delivered Meterset	(3008,0044)	DS	1	
>>Nominal Beam Energy	(300A,0114)	DS	1C	
>>Ion Wedge Position Sequence	(300A,03AC)	SQ	1C	
>>>Referenced Wedge Number	(300C,00C0)	IS	1	

>>>Wedge Position	(300A,0118)	CS	1	
>>>Wedge Thin Edge Position	(300A,00DB)	FL	1C	
>>Lateral Spreading Device Settings Sequence	(300A,0370)	SQ	1C	
>>>Referenced Lateral Spreading Device Number	(300C,0102)	IS	1	
>>>Lateral Spreading Device Setting	(300A,0372)	LO	1	
>>Range Modulator Settings Sequence	(300A,0380)	SQ	1C	
>>>Referenced Range Modulator Number	(300C,0104)	IS	1	
>>>Range Modulator Gating Start Value	(300A,0382)	FL	1C	
>>>Range Modulator Gating Stop Value	(300A,0384)	FL	1C	
>>Gantry Angle	(300A,011E)	DS	1C	
>>Gantry Rotation Direction	(300A,011F)	CS	1C	
>>Beam Limiting Device Angle	(300A,0120)	DS	1C	
>>Beam Limiting Device Rotation Direction	(300A,0121)	CS	1C	
>>Scan Spot Tune ID	(300A,0390)	SH	1C	
>>Number of Scan Spot Positions	(300A,0392)	IS	1C	
>>Scan Spot Position Map	(300A,0394)	FL	1C	
>>Scan Spot Metersets Delivered	(3008,0047)	FL	1C	
>>Number of Paintings	(300A,039A)	IS	1C	
>>Patient Support Angle	(300A,0122)	DS	1C	
>>Patient Support Rotation Direction	(300A,0123)	CS	1C	
>>Table Top Pitch Angle	(300A,0140)	FL	2C	
>>Table Top Pitch Rotation Direction	(300A,0142)	CS	2C	
>>Table Top Roll Angle	(300A,0144)	FL	2C	
>>Table Top Roll Rotation Direction	(300A,0146)	CS	2C	
>>Table Top Vertical Position	(300A,0128)	DS	2C	
>>Table Top Longitudinal Position	(300A,0129)	DS	2C	
>>Table Top Lateral Position	(300A,012A)	DS	2C	
>>Snout Position	(300A,030D)	FL	2C	

9.1.1.2.8 SOP Common Module

Attribute name	Tag	Vr	Type	Comment
SOP Class UID	(0008,0016)	UI	1	
SOP Instance UID	(0008,0018)	UI	1	

9.1.1.3 RT Beams Delivery Instruction IOD

IE	Module	Used
Patient	Patient Module	Yes
Study	General Study Module	Yes
Series	General Series Module	Yes
Equipment	General Equipment Module	Yes
Plan	RT Beams Delivery Instruction Module	Yes
	Common Instance Reference Module	No
	SOP Common Module	Yes

9.1.1.3.1 Patient Module

Attribute name	Tag	Vr	Type	Comment
Patient's Name	(0010,0010)	PN	2	
Patient ID	(0010,0020)	LO	2	
Patient's Birth Date	(0010,0030)	DA	2	
Patient's Sex	(0010,0040)	CS	2	Possible values: M, F, O.

9.1.1.3.2 General Study Module

Attribute name	Tag	Vr	Type	Comment
Study Instance UID	(0020,000D)	UI	1	
Study Date	(0008,0020)	DA	2	
Study Time	(0008,0030)	TM	2	
Referring Physician's Name	(0008,0090)	PN	2	
Study ID	(0020,0010)	SH	2	
Accession Number	(0008,0050)	SH	2	

9.1.1.3.3 General Series Module

Attribute name	Tag	Vr	Type	Comment
Modality	(0008,0060)	CS	1	Always PLAN.
Series Instance UID	(0020,000E)	UI	1	
Series Number	(0020,0011)	IS	2	

9.1.1.3.4 General Equipment Module

Attribute name	Tag	Vr	Type	Comment
Manufacturer	(0008,0070)	LO	2	Always RaySearch Laboratories.

9.1.1.3.5 RT Beams Delivery Instruction Module

Attribute name	Tag	Vr	Type	Comment
Referenced RT Plan Sequence	(300C,0002)	SQ	1	
>Referenced SOP Class UID	(0008,1150)	UI	1	
>Referenced SOP Instance UID	(0008,1155)	UI	1	
Beam Task Sequence	(0074,1020)	SQ	1	
>Beam Task Type	(0074,1022)	CS	1	Always TREAT.
>Treatment Delivery Type	(300A,00CE)	CS	1	Possible values: TREATMENT, CONTINUATION.
>Continuation Start Meterset	(0074,0120)	FD	1C	
>Continuation End Meterset	(0074,0121)	FD	1C	
>Current Fraction Number	(3008,0022)	IS	1	
>Referenced Beam Number	(300C,0006)	IS	1	
>Table Top Vertical Adjusted Position	(0074,1026)	FD	2	
>Table Top Longitudinal Adjusted Position	(0074,1027)	FD	2	
>Table Top Lateral Adjusted Position	(0074,1028)	FD	2	
>Patient Support Adjusted Angle	(0074,102A)	FD	2	
>Table Top Pitch Adjusted Angle	(0074,102C)	FD	2	
>Table Top Roll Adjusted Angle	(0074,102D)	FD	2	
Omitted Beam Task Sequence	(300C,0111)	SQ	3	
>Referenced Beam Number	(300C,0006)	IS	1	
>Reason for Omission	(300C,0112)	CS	1	Always ALREADY_TREATED.

9.1.1.3.6 SOP Common Module

Attribute name	Tag	Vr	Type	Comment
SOP Class UID	(0008,0016)	UI	1	
SOP Instance UID	(0008,0018)	UI	1	

9.1.1.4 Unified Procedure Step IOD

IE	Module	Used
Unified Procedure Step	SOP Common Module	Yes
	Unified Procedure Step Relationship Module	Yes
	Unified Procedure Step Scheduled Procedure Information Module	Yes
	Unified Procedure Step Progress Information Module	Yes
	Unified Procedure Step Performed Procedure Information Module	No
	Patient Demographic Module	No
	Patient Medical Module	No
	Visit Identification Module	No
	Visit Status Module	No
	Visit Admission Module	No

9.1.1.4.1 SOP Common Module

Attribute name	Tag	Vr	Type	Comment
SOP Class UID	(0008,0016)	UI	1	
SOP Instance UID	(0008,0018)	UI	1	
Timezone Offset From UTC	(0008,0201)	SH	3	

9.1.1.4.2 Unified Procedure Step Relationship Module

Attribute name	Tag	Vr	Type	Comment
Patient's Name	(0010,0010)	PN		Patient name
Patient ID	(0010,0020)	LO		Patient ID
Patient's Birth Date	(0010,0030)	DA		Patient birth date
Patient's Sex	(0010,0040)	CS		Patient sex Possible values: M, F, O.

9.1.1.4.3 Unified Procedure Step Scheduled Procedure Information Module

Attribute name	Tag	Vr	Type	Comment
Scheduled Procedure Step Priority	(0074,1200)	CS		Always MEDIUM.
Procedure Step Label	(0074,1204)	LO		
Scheduled Station Name Code Sequence	(0040,4025)	SQ		
>Code Value	(0008,0100)	SH	1C	Scheduled station name
>Coding Scheme Designator	(0008,0102)	SH	1C	If included in request, otherwise RAYSEARCH
>Code Meaning	(0008,0104)	LO	1	
Scheduled Procedure Step Start DateTime	(0040,4005)	DT		
Expected Completion DateTime	(0040,4011)	DT		
Scheduled Workitem Code Sequence	(0040,4018)	SQ		
>Code Value	(0008,0100)	SH	1C	Always <ul style="list-style-type: none"> 121726 - RT Treatment with Internal Verification
>Coding Scheme Designator	(0008,0102)	SH	1C	Always DCM.

>Code Meaning	(0008,0104)	LO	1	Always RT Treatment with Internal Verification.
Scheduled Processing Parameters Sequence	(0074,1210)	SQ		
>Value Type	(0040,A040)	CS	1	Possible values: TEXT, NUMERIC.
>Concept Name Code Sequence	(0040,A043)	SQ	1	
>>Code Value	(0008,0100)	SH	1C	Always 2008001.
>>Coding Scheme Designator	(0008,0102)	SH	1C	Always 99IHER02008.
>>Code Meaning	(0008,0104)	LO	1	Always Treatment Delivery Type.
>Text Value	(0040,A160)	UT	1C	Possible values: <ul style="list-style-type: none"> • CONTINUATION - For continuation fractions. • TREATMENT - For standard fractions.
>Numeric Value	(0040,A30A)	DS	1C	
>Measurement Units Code Sequence	(0040,08EA)	SQ	1C	
>>Code Value	(0008,0100)	SH	1C	Always 1.
>>Coding Scheme Designator	(0008,0102)	SH	1C	Always UCUM.
>>Code Meaning	(0008,0104)	LO	1	Always No Units.
Input Readiness State	(0040,4041)	CS		Possible values: <ul style="list-style-type: none"> • INCOMPLETE - The Input Information Sequence is not yet complete and additional instance references might be added. • UNAVAILABLE - The Input Information Sequence is complete but one or more of the referenced instances might not yet be available from the referenced source(s). • READY - The Input Information Sequence is complete and the referenced instances are available from the referenced sources.
Input Information Sequence	(0040,4021)	SQ		
>Type of Instances	(0040,E020)	CS	1	Always DICOM.
>Study Instance UID	(0020,0000)	UI	1C	
>Series Instance UID	(0020,000E)	UI	1C	
>Referenced SOP Sequence	(0008,1199)	SQ	1	
>>Referenced SOP Class UID	(0008,1150)	UI	1	
>>Referenced SOP Instance UID	(0008,1155)	UI	1	
>DICOM Retrieval Sequence	(0040,E021)	SQ	1C	
>>Retrieve AE Title	(0008,0054)	AE	1	
Study Instance UID	(0020,0000)	UI		

9.1.1.4.4 Unified Procedure Step Progress Information Module

Attribute name	Tag	Vr	Type	Comment
Procedure Step State	(0074,1000)	CS		Possible values: SCHEDULED, IN PROGRESS, CANCELED, COMPLETED.

9.1.2 Usage of Attributes From Received IODs

9.1.2.1 CT Image IOD

IE	Module	Used
Patient	Patient Module	Yes
Study	General Study Module	Yes
Series	General Series Module	Yes
Frame of Reference	Frame of Reference Module	Yes
Equipment	General Equipment Module	No

Image	General Image Module	No
	Image Plane Module	No
	Image Pixel Module	No
	Contrast/Bolus Module	No
	CT Image Module	Yes
	Multi-energy CT Image Module	No
	SOP Common Module	Yes

9.1.2.1.1 Patient Module

Attribute name	Tag	Vr	Type	Comment
Patient's Name	(0010,0010)	PN	2	
Patient ID	(0010,0020)	LO	2	
Patient's Birth Date	(0010,0030)	DA	2	
Patient's Sex	(0010,0040)	CS	2	

9.1.2.1.2 General Study Module

Attribute name	Tag	Vr	Type	Comment
Study Instance UID	(0020,000D)	UI	1	

9.1.2.1.3 General Series Module

Attribute name	Tag	Vr	Type	Comment
Modality	(0008,0060)	CS	1	Value not read
Series Instance UID	(0020,000E)	UI	1	
Series Date	(0008,0021)	DA	3	
Series Time	(0008,0031)	TM	3	
Series Description	(0008,103E)	LO	3	
Patient Position	(0018,5100)	CS	2C	

9.1.2.1.4 Frame of Reference Module

Attribute name	Tag	Vr	Type	Comment
Frame of Reference UID	(0020,0052)	UI	1	

9.1.2.1.5 CT Image Module

Attribute name	Tag	Vr	Type	Comment
Image Type	(0008,0008)	CS	1	Supported values: ORIGINAL, PRIMARY, AXIAL, CBCT.
Samples per Pixel	(0028,0002)	US	1	Value not read
Photometric Interpretation	(0028,0004)	CS	1	Value not read
Bits Allocated	(0028,0100)	US	1	Value not read
Bits Stored	(0028,0101)	US	1	Value not read
High Bit	(0028,0102)	US	1	Value not read
Rescale Intercept	(0028,1052)	DS	1	Value not read
Rescale Slope	(0028,1053)	DS	1	Value not read
Table Height	(0018,1130)	DS	3	
Patient Support Angle	(300A,0122)	DS	3	
Table Top Pitch Angle	(300A,0140)	FL	3	
Table Top Roll Angle	(300A,0144)	FL	3	
Table Top Longitudinal Position	(300A,0129)	DS	3	

Table Top Lateral Position	(300A,012A)	DS	3	
----------------------------	-------------	----	---	--

9.1.2.1.6 SOP Common Module

Attribute name	Tag	Vr	Type	Comment
SOP Class UID	(0008,0016)	UI	1	
SOP Instance UID	(0008,0018)	UI	1	

9.1.2.2 RT Image IOD

IE	Module	Used
Patient	Patient Module	Yes
Study	General Study Module	Yes
Series	RT Series Module	Yes
Frame of Reference	Frame of Reference Module	Yes
Equipment	General Equipment Module	No
Image	General Image Module	No
	Image Pixel Module	No
	Contrast/Bolus Module	No
	Cine Module	No
	Multi-frame Module	No
	RT Image Module	Yes
	SOP Common Module	Yes
	Frame Extraction Module	No

9.1.2.2.1 Patient Module

Attribute name	Tag	Vr	Type	Comment
Patient's Name	(0010,0010)	PN	2	
Patient ID	(0010,0020)	LO	2	
Patient's Birth Date	(0010,0030)	DA	2	
Patient's Sex	(0010,0040)	CS	2	

9.1.2.2.2 General Study Module

Attribute name	Tag	Vr	Type	Comment
Study Instance UID	(0020,000D)	UI	1	

9.1.2.2.3 RT Series Module

Attribute name	Tag	Vr	Type	Comment
Modality	(0008,0060)	CS	1	Value not read
Series Instance UID	(0020,000E)	UI	1	
Series Date	(0008,0021)	DA	3	
Series Time	(0008,0031)	TM	3	

9.1.2.2.4 Frame of Reference Module

Attribute name	Tag	Vr	Type	Comment
Frame of Reference UID	(0020,0052)	UI	1	

9.1.2.2.5 RT Image Module

Attribute name	Tag	Vr	Type	Comment
Samples per Pixel	(0028,0002)	US	1	Value not read

Photometric Interpretation	(0028,0004)	CS	1	Value not read
Bits Allocated	(0028,0100)	US	1	Value not read
Bits Stored	(0028,0101)	US	1	Value not read
High Bit	(0028,0102)	US	1	Value not read
Pixel Representation	(0028,0103)	US	1	Value not read
RT Image Label	(3002,0002)	SH	1	
Image Type	(0008,0008)	CS	1	Supported values: DERIVED, SECONDARY, DRR, CT_PROJECTION, PORTAL.
RT Image Plane	(3002,000C)	CS	1	Value not read
Patient Support Angle	(300A,0122)	DS	3	
Table Top Pitch Angle	(300A,0140)	FL	3	
Table Top Roll Angle	(300A,0144)	FL	3	
Table Top Vertical Position	(300A,0128)	DS	3	
Table Top Longitudinal Position	(300A,0129)	DS	3	
Table Top Lateral Position	(300A,012A)	DS	3	
Patient Position	(0018,5100)	CS	1C	

9.1.2.2.6 SOP Common Module

Attribute name	Tag	Vr	Type	Comment
SOP Class UID	(0008,0016)	UI	1	
SOP Instance UID	(0008,0018)	UI	1	

9.1.2.3 RT Structure Set IOD

IE	Module	Used
Patient	Patient Module	Yes
Study	General Study Module	Yes
Series	RT Series Module	Yes
Equipment	General Equipment Module	No
Structure Set	Structure Set Module	Yes
	ROI Contour Module	No
	RT ROI Observations Module	No
	SOP Common Module	Yes

9.1.2.3.1 Patient Module

Attribute name	Tag	Vr	Type	Comment
Patient's Name	(0010,0010)	PN	2	
Patient ID	(0010,0020)	LO	2	
Patient's Birth Date	(0010,0030)	DA	2	
Patient's Sex	(0010,0040)	CS	2	

9.1.2.3.2 General Study Module

Attribute name	Tag	Vr	Type	Comment
Study Instance UID	(0020,000D)	UI	1	

9.1.2.3.3 RT Series Module

Attribute name	Tag	Vr	Type	Comment
Modality	(0008,0060)	CS	1	Value not read
Series Instance UID	(0020,000E)	UI	1	

9.1.2.3.4 Structure Set Module

Attribute name	Tag	Vr	Type	Comment
Structure Set Label	(3006,0002)	SH	1	Value not read
Referenced Frame of Reference Sequence	(3006,0010)	SQ	3	
>Frame of Reference UID	(0020,0052)	UI	1	
>RT Referenced Study Sequence	(3006,0012)	SQ	3	
>>Referenced SOP Class UID	(0008,1150)	UI	1	Value not read
>>Referenced SOP Instance UID	(0008,1155)	UI	1	Value not read
>>RT Referenced Series Sequence	(3006,0014)	SQ	1	
>>>Series Instance UID	(0020,000E)	UI	1	
>>>Contour Image Sequence	(3006,0016)	SQ	1	
>>>>Referenced SOP Class UID	(0008,1150)	UI	1	
>>>>Referenced SOP Instance UID	(0008,1155)	UI	1	
Structure Set ROI Sequence	(3006,0020)	SQ	1	Value not read

9.1.2.3.5 SOP Common Module

Attribute name	Tag	Vr	Type	Comment
SOP Class UID	(0008,0016)	UI	1	
SOP Instance UID	(0008,0018)	UI	1	

9.1.2.4 Spatial Registration IOD

IE	Module	Used
Patient	Patient Module	Yes
Study	General Study Module	Yes
Series	General Series Module	Yes
	Spatial Registration Series Module	No
Frame of Reference	Frame of Reference Module	Yes
Equipment	General Equipment Module	No
Spatial Registration	Spatial Registration Module	Yes
	Common Instance Reference Module	Yes
	SOP Common Module	Yes

9.1.2.4.1 Patient Module

Attribute name	Tag	Vr	Type	Comment
Patient's Name	(0010,0010)	PN	2	
Patient ID	(0010,0020)	LO	2	
Patient's Birth Date	(0010,0030)	DA	2	
Patient's Sex	(0010,0040)	CS	2	

9.1.2.4.2 General Study Module

Attribute name	Tag	Vr	Type	Comment
Study Instance UID	(0020,000D)	UI	1	

9.1.2.4.3 General Series Module

Attribute name	Tag	Vr	Type	Comment
Modality	(0008,0060)	CS	1	Value not read
Series Instance UID	(0020,000E)	UI	1	

9.1.2.4.4 Frame of Reference Module

Attribute name	Tag	Vr	Type	Comment
Frame of Reference UID	(0020,0052)	UI	1	

9.1.2.4.5 Spatial Registration Module

Attribute name	Tag	Vr	Type	Comment
Content Date	(0008,0023)	DA	1	Value not read
Content Time	(0008,0033)	TM	1	Value not read
Instance Number	(0020,0013)	IS	1	Value not read
Content Label	(0070,0080)	CS	1	Value not read
Registration Sequence	(0070,0308)	SQ	1	
>Frame of Reference UID	(0020,0052)	UI	1C	
>Referenced Image Sequence	(0008,1140)	SQ	1C	
>>Referenced SOP Class UID	(0008,1150)	UI	1	
>>Referenced SOP Instance UID	(0008,1155)	UI	1	
>Matrix Registration Sequence	(0070,0309)	SQ	1	Value not read

9.1.2.4.6 Common Instance Reference Module

Attribute name	Tag	Vr	Type	Comment
Referenced Series Sequence	(0008,1115)	SQ	1C	
>Series Instance UID	(0020,000E)	UI	1	
>Referenced Instance Sequence	(0008,114A)	SQ	1	
>>Referenced SOP Class UID	(0008,1150)	UI	1	
>>Referenced SOP Instance UID	(0008,1155)	UI	1	

9.1.2.4.7 SOP Common Module

Attribute name	Tag	Vr	Type	Comment
SOP Class UID	(0008,0016)	UI	1	
SOP Instance UID	(0008,0018)	UI	1	

9.1.2.5 RT Ion Plan IOD

IE	Module	Used
Patient	Patient Module	Yes
Study	General Study Module	Yes
Series	RT Series Module	Yes
Frame of Reference	Frame of Reference Module	No
Equipment	General Equipment Module	Yes
Plan	RT General Plan Module	Yes
	RT Patient Setup Module	Yes
	RT Fraction Scheme Module	Yes
	RT Ion Beams Module	Yes
	SOP Common Module	Yes

9.1.2.5.1 Patient Module

Attribute name	Tag	Vr	Type	Comment
Patient's Name	(0010,0010)	PN	2	
Patient ID	(0010,0020)	LO	2	

Patient's Birth Date	(0010,0030)	DA	2	
Patient's Sex	(0010,0040)	CS	2	

9.1.2.5.2 General Study Module

Attribute name	Tag	Vr	Type	Comment
Study Instance UID	(0020,000D)	UI	1	
Study Date	(0008,0020)	DA	2	
Study Time	(0008,0030)	TM	2	
Referring Physician's Name	(0008,0090)	PN	2	
Study ID	(0020,0010)	SH	2	
Accession Number	(0008,0050)	SH	2	
Study Description	(0008,1030)	LO	3	

9.1.2.5.3 RT Series Module

Attribute name	Tag	Vr	Type	Comment
Modality	(0008,0060)	CS	1	Value not read
Series Instance UID	(0020,000E)	UI	1	

9.1.2.5.4 General Equipment Module

Attribute name	Tag	Vr	Type	Comment
Software Versions	(0018,1020)	LO	3	Read from the plan to determine if checksum validation shall be run.

9.1.2.5.5 RT General Plan Module

Attribute name	Tag	Vr	Type	Comment
RT Plan Label	(300A,0002)	SH	1	
RT Plan Geometry	(300A,000C)	CS	1	Value not read

9.1.2.5.6 RT Patient Setup Module

Attribute name	Tag	Vr	Type	Comment
Patient Setup Sequence	(300A,0180)	SQ	1	
>Patient Setup Number	(300A,0182)	IS	1	
>Patient Position	(0018,5100)	CS	1C	Supported value: HFS.
>Table Top Vertical Setup Displacement	(300A,01D2)	DS	3	
>Table Top Longitudinal Setup Displacement	(300A,01D4)	DS	3	
>Table Top Lateral Setup Displacement	(300A,01D6)	DS	3	

9.1.2.5.7 RT Fraction Scheme Module

Attribute name	Tag	Vr	Type	Comment
Fraction Group Sequence	(300A,0070)	SQ	1	
>Fraction Group Number	(300A,0071)	IS	1	Value not read
>Number of Fractions Planned	(300A,0078)	IS	2	
>Number of Beams	(300A,0080)	IS	1	Value not read
>Number of Brachy Application Setups	(300A,00A0)	IS	1	Value not read

9.1.2.5.8 RT Ion Beams Module

Attribute name	Tag	Vr	Type	Comment
Ion Beam Sequence	(300A,03A2)	SQ	1	
>Beam Number	(300A,00C0)	IS	1	
>Beam Name	(300A,00C2)	LO	1	

>Beam Type	(300A,00C4)	CS	1	
>Radiation Type	(300A,00C6)	CS	1	
>Scan Mode	(300A,0308)	CS	1	
>Treatment Machine Name	(300A,00B2)	SH	2	
>Primary Dosimeter Unit	(300A,00B3)	CS	1	Value not read
>Virtual Source-Axis Distances	(300A,030A)	FL	1	Value not read
>Referenced Patient Setup Number	(300C,006A)	IS	3	
>Treatment Delivery Type	(300A,00CE)	CS	1	Supported values: TREATMENT, SETUP.
>Number of Wedges	(300A,00D0)	IS	1	
>Number of Compensators	(300A,00E0)	IS	1	
>Ion Range Compensator Sequence	(300A,02EA)	SQ	1C	
>>Compensator Number	(300A,00E4)	IS	1	
>>Compensator ID	(300A,00E5)	SH	3	
>>Compensator Divergence	(300A,02E0)	CS	1	Value not read
>>Compensator Mounting Position	(300A,02E1)	CS	1	Value not read
>>Compensator Rows	(300A,00E7)	IS	1	Value not read
>>Compensator Columns	(300A,00E8)	IS	1	Value not read
>>Compensator Pixel Spacing	(300A,00E9)	DS	1	Value not read
>>Compensator Position	(300A,00EA)	DS	1	Value not read
>>Compensator Thickness Data	(300A,00EC)	DS	1	Value not read
>Number of Boli	(300A,00ED)	IS	1	
>Referenced Bolus Sequence	(300C,00B0)	SQ	1C	
>>Referenced ROI Number	(3006,0084)	IS	1	
>Number of Blocks	(300A,00F0)	IS	1	
>Ion Block Sequence	(300A,03A6)	SQ	1C	
>>Isocenter to Block Tray Distance	(300A,00F7)	FL	1	Value not read
>>Block Type	(300A,00F8)	CS	1	Value not read
>>Block Divergence	(300A,00FA)	CS	1	Value not read
>>Block Mounting Position	(300A,00FB)	CS	1	Value not read
>>Block Number	(300A,00FC)	IS	1	
>>Block Thickness	(300A,0100)	DS	1	Value not read
>>Block Number of Points	(300A,0104)	IS	1	Value not read
>>Block Data	(300A,0106)	DS	1	Value not read
>Snout Sequence	(300A,030C)	SQ	3	
>>Snout ID	(300A,030F)	SH	1	
>Applicator Sequence	(300A,0107)	SQ	3	
>>Applicator ID	(300A,0108)	SH	1	
>>Applicator Type	(300A,0109)	CS	1	
>Number of Range Shifters	(300A,0312)	IS	1	
>Range Shifter Sequence	(300A,0314)	SQ	1C	
>>Range Shifter Number	(300A,0316)	IS	1	
>>Range Shifter ID	(300A,0318)	SH	1	
>>Range Shifter Type	(300A,0320)	CS	1	Value not read
>Number of Lateral Spreading Devices	(300A,0330)	IS	1	

>Lateral Spreading Device Sequence	(300A,0332)	SQ	1C	
>>Lateral Spreading Device Number	(300A,0334)	IS	1	
>>Lateral Spreading Device ID	(300A,0336)	SH	1	
>>Lateral Spreading Device Type	(300A,0338)	CS	1	
>Number of Range Modulators	(300A,0340)	IS	1	
>Range Modulator Sequence	(300A,0342)	SQ	1C	
>>Range Modulator Number	(300A,0344)	IS	1	
>>Range Modulator ID	(300A,0346)	SH	1	
>>Range Modulator Type	(300A,0348)	CS	1	
>>Beam Current Modulation ID	(300A,034C)	SH	1C	
>Patient Support Type	(300A,0350)	CS	1	
>Number of Control Points	(300A,0110)	IS	1	Value not read
>Ion Control Point Sequence	(300A,03A8)	SQ	1	
>>Control Point Index	(300A,0112)	IS	1	
>>Nominal Beam Energy	(300A,0114)	DS	1C	
>>Ion Wedge Position Sequence	(300A,03AC)	SQ	1C	
>>>Referenced Wedge Number	(300C,00C0)	IS	1	
>>>Wedge Position	(300A,0118)	CS	1	
>>>Wedge Thin Edge Position	(300A,00DB)	FL	1C	
>>Lateral Spreading Device Settings Sequence	(300A,0370)	SQ	1C	
>>>Referenced Lateral Spreading Device Number	(300C,0102)	IS	1	
>>>Lateral Spreading Device Setting	(300A,0372)	LO	1	
>>Range Modulator Settings Sequence	(300A,0380)	SQ	1C	
>>>Referenced Range Modulator Number	(300C,0104)	IS	1	
>>>Range Modulator Gating Start Value	(300A,0382)	FL	1C	
>>>Range Modulator Gating Stop Value	(300A,0384)	FL	1C	
>>Gantry Angle	(300A,011E)	DS	1C	
>>Gantry Rotation Direction	(300A,011F)	CS	1C	
>>Beam Limiting Device Angle	(300A,0120)	DS	1C	
>>Beam Limiting Device Rotation Direction	(300A,0121)	CS	1C	
>>Scan Spot Tune ID	(300A,0390)	SH	1C	
>>Number of Scan Spot Positions	(300A,0392)	IS	1C	
>>Scan Spot Position Map	(300A,0394)	FL	1C	
>>Number of Paintings	(300A,039A)	IS	1C	
>>Patient Support Angle	(300A,0122)	DS	1C	
>>Patient Support Rotation Direction	(300A,0123)	CS	1C	
>>Table Top Pitch Angle	(300A,0140)	FL	2C	
>>Table Top Pitch Rotation Direction	(300A,0142)	CS	2C	
>>Table Top Roll Angle	(300A,0144)	FL	2C	
>>Table Top Roll Rotation Direction	(300A,0146)	CS	2C	
>>Table Top Vertical Position	(300A,0128)	DS	2C	
>>Table Top Longitudinal Position	(300A,0129)	DS	2C	

>>Table Top Lateral Position	(300A,012A)	DS	2C	
>>Snout Position	(300A,030D)	FL	2C	
>>Isocenter Position	(300A,012C)	DS	2C	
>RaySearch Private Creator	(4001,0010)	LO	3	RAYSEARCHLABS 2.0
>Internal Treatment Machine Name	(4001,1012)	SH	3	RaySearch Private tag. The internal treatment machine name. This value will differ from Treatment Machine Name (300A,00B2) if a treatment machine name alias have been specified on the ion beam quality.

9.1.2.5.9 SOP Common Module

Attribute name	Tag	Vr	Type	Comment
SOP Class UID	(0008,0016)	UI	1	
SOP Instance UID	(0008,0018)	UI	1	
RaySearch Private Creator	(4001,0010)	LO	3	RAYSEARCHLABS 2.0
RaySearch Checksum Algorithm Version	(4001,1060)	LO	3	RaySearch checksum algorithm version used to calculate the checksum of the plan.
RaySearch Checksum Data	(4001,1061)	OB	3	RaySearch custom checksum calculation specific for the current checksum algorithm version.

9.1.2.6 RT Ion Beams Treatment Record IOD

IE	Module	Used
Patient	Patient Module	Yes
Study	General Study Module	Yes
Series	RT Series Module	Yes
Equipment	General Equipment Module	No
Treatment Record	RT General Treatment Record Module	Yes
	RT Treatment Machine Record Module	Yes
	RT Ion Beams Session Record Module	Yes
	SOP Common Module	Yes
	Common Instance Reference Module	Yes

9.1.2.6.1 Patient Module

Attribute name	Tag	Vr	Type	Comment
Patient's Name	(0010,0010)	PN	2	
Patient ID	(0010,0020)	LO	2	
Patient's Birth Date	(0010,0030)	DA	2	
Patient's Sex	(0010,0040)	CS	2	

9.1.2.6.2 General Study Module

Attribute name	Tag	Vr	Type	Comment
Study Instance UID	(0020,000D)	UI	1	

9.1.2.6.3 RT Series Module

Attribute name	Tag	Vr	Type	Comment
Modality	(0008,0060)	CS	1	Value not read
Series Instance UID	(0020,000E)	UI	1	

9.1.2.6.4 RT General Treatment Record Module

Attribute name	Tag	Vr	Type	Comment
Instance Number	(0020,0013)	IS	1	Value not read

Treatment Date	(3008,0250)	DA	2	
Treatment Time	(3008,0251)	TM	2	
Referenced RT Plan Sequence	(300C,0002)	SQ	2	
>Referenced SOP Class UID	(0008,1150)	UI	1	
>Referenced SOP Instance UID	(0008,1155)	UI	1	

9.1.2.6.5 RT Treatment Machine Record Module

Attribute name	Tag	Vr	Type	Comment
Treatment Machine Sequence	(300A,0206)	SQ	1	
>Treatment Machine Name	(300A,00B2)	SH	2	Used for validation purposes when reading treatment record from disk.

9.1.2.6.6 RT Ion Beams Session Record Module

Attribute name	Tag	Vr	Type	Comment
Primary Dosimeter Unit	(300A,00B3)	CS	1	
Treatment Session Ion Beam Sequence	(3008,0021)	SQ	1	
>Referenced Beam Number	(300C,0006)	IS	1	
>Beam Name	(300A,00C2)	LO	1	
>Beam Type	(300A,00C4)	CS	1	Value not read
>Radiation Type	(300A,00C6)	CS	1	Value not read
>Scan Mode	(300A,0308)	CS	1	Value not read
>Number of Wedges	(300A,00D0)	IS	1	Value not read
>Number of Compensators	(300A,00E0)	IS	1	Value not read
>Number of Boli	(300A,00ED)	IS	1	Value not read
>Number of Blocks	(300A,00F0)	IS	1	Value not read
>Recorded Snout Sequence	(3008,00F0)	SQ	1C	
>>Snout ID	(300A,030F)	SH	1	
>Number of Range Shifters	(300A,0312)	IS	1	Value not read
>Number of Lateral Spreading Devices	(300A,0330)	IS	1	Value not read
>Number of Range Modulators	(300A,0340)	IS	1	Value not read
>Patient Support Type	(300A,0350)	CS	1	Value not read
>Current Fraction Number	(3008,0022)	IS	2	
>Treatment Delivery Type	(300A,00CE)	CS	2	Supported values: TREATMENT, CONTINUATION.
>Treatment Termination Status	(3008,002A)	CS	1	Supported values: NORMAL, OPERATOR, MACHINE, UNKNOWN.
>Specified Primary Meterset	(3008,0032)	DS	3	
>Specified Secondary Meterset	(3008,0033)	DS	3	
>Delivered Primary Meterset	(3008,0036)	DS	3	
>Delivered Secondary Meterset	(3008,0037)	DS	3	
>Specified Treatment Time	(3008,003A)	DS	3	
>Delivered Treatment Time	(3008,003B)	DS	3	
>Number of Control Points	(300A,0110)	IS	1	Value not read
>Ion Control Point Delivery Sequence	(3008,0041)	SQ	1	
>>Referenced Control Point Index	(300C,00F0)	IS	1	Value not read
>>Treatment Control Point Date	(3008,0024)	DA	1	Value not read
>>Treatment Control Point Time	(3008,0025)	TM	1	Value not read
>>Delivered Meterset	(3008,0044)	DS	1	Value not read

>>Gantry Angle	(300A,011E)	DS	1C	
>>Patient Support Angle	(300A,0122)	DS	1C	
>>Table Top Pitch Angle	(300A,0140)	FL	2C	
>>Table Top Roll Angle	(300A,0144)	FL	2C	
>>Table Top Vertical Position	(300A,0128)	DS	2C	
>>Table Top Longitudinal Position	(300A,0129)	DS	2C	
>>Table Top Lateral Position	(300A,012A)	DS	2C	
>>Snout Position	(300A,030D)	FL	2C	
ProNova Solutions	(0009,00FF)	LO	3	ProNova Solutions
ProNovaPositioningUser	(0009,FF00)	LO	3	
ProNovaSetupNotes	(0009,FF01)	LO	3	
ProNovaClosingNotes	(0009,FF02)	LO	3	

9.1.2.6.7 SOP Common Module

Attribute name	Tag	Vr	Type	Comment
SOP Class UID	(0008,0016)	UI	1	
SOP Instance UID	(0008,0018)	UI	1	

9.1.2.6.8 Common Instance Reference Module

Attribute name	Tag	Vr	Type	Comment
Referenced Series Sequence	(0008,1115)	SQ	1C	
>Series Instance UID	(0020,000E)	UI	1	Used to find the series of the Referenced RT Plan (300C,0002).
>Referenced Instance Sequence	(0008,114A)	SQ	1	
>>Referenced SOP Class UID	(0008,1150)	UI	1	Used to find the series of the Referenced RT Plan (300C,0002).
>>Referenced SOP Instance UID	(0008,1155)	UI	1	Used to find the series of the Referenced RT Plan (300C,0002).

9.1.2.7 RT Beams Delivery Instruction IO

IE	Module	Used
Patient	Patient Module	Yes
Study	General Study Module	Yes
Series	General Series Module	Yes
Equipment	General Equipment Module	No
Plan	RT Beams Delivery Instruction Module	Yes
	Common Instance Reference Module	No
	SOP Common Module	Yes

9.1.2.7.1 Patient Module

Attribute name	Tag	Vr	Type	Comment
----------------	-----	----	------	---------

9.1.2.7.2 General Study Module

Attribute name	Tag	Vr	Type	Comment
Study Instance UID	(0020,000D)	UI	1	

9.1.2.7.3 General Series Module

Attribute name	Tag	Vr	Type	Comment
Modality	(0008,0060)	CS	1	Value not read
Series Instance UID	(0020,000E)	UI	1	

9.1.2.7.4 RT Beams Delivery Instruction Module

Attribute name	Tag	Vr	Type	Comment
Referenced RT Plan Sequence	(300C,0002)	SQ	1	Value not read
Beam Task Sequence	(0074,1020)	SQ	1	
>Beam Task Type	(0074,1022)	CS	1	Value not read
>Treatment Delivery Type	(300A,00CE)	CS	1	Supported values: TREATMENT, CONTINUATION.
>Continuation Start Meterset	(0074,0120)	FD	1C	
>Continuation End Meterset	(0074,0121)	FD	1C	
>Current Fraction Number	(3008,0022)	IS	1	Value not read
>Referenced Beam Number	(300C,0006)	IS	1	
Omitted Beam Task Sequence	(300C,0111)	SQ	3	
>Referenced Beam Number	(300C,0006)	IS	1	
>Reason for Omission	(300C,0112)	CS	1	Value not read

9.1.2.7.5 SOP Common Module

Attribute name	Tag	Vr	Type	Comment
SOP Class UID	(0008,0016)	UI	1	
SOP Instance UID	(0008,0018)	UI	1	

9.1.2.8 Unified Procedure Step IOB

IE	Module	Used
Unified Procedure Step	SOP Common Module	Yes
	Unified Procedure Step Relationship Module	Yes
	Unified Procedure Step Scheduled Procedure Information Module	Yes
	Unified Procedure Step Progress Information Module	Yes
	Unified Procedure Step Performed Procedure Information Module	No
	Patient Demographic Module	No
	Patient Medical Module	No
	Visit Identification Module	No
	Visit Status Module	No
	Visit Admission Module	No

9.1.2.8.1 SOP Common Module

Attribute name	Tag	Vr	Type	Comment
SOP Class UID	(0008,0016)	UI	1	Value not read
SOP Instance UID	(0008,0018)	UI	1	Value not read
Timezone Offset From UTC	(0008,0201)	SH	3	

9.1.2.8.2 Unified Procedure Step Relationship Module

Attribute name	Tag	Vr	Type	Comment
Patient's Name	(0010,0010)	PN		
Patient ID	(0010,0020)	LO		

9.1.2.8.3 Unified Procedure Step Scheduled Procedure Information Module

Attribute name	Tag	Vr	Type	Comment
Scheduled Station Name Code Sequence	(0040,4025)	SQ		

>Code Value	(0008,0100)	SH	1C	
>Coding Scheme Designator	(0008,0102)	SH	1C	
>Code Meaning	(0008,0104)	LO	1	
Scheduled Workitem Code Sequence	(0040,4018)	SQ		
>Code Value	(0008,0100)	SH	1C	
>Coding Scheme Designator	(0008,0102)	SH	1C	
>Code Meaning	(0008,0104)	LO	1	

9.1.2.8.4 Unified Procedure Step Progress Information Module

Attribute name	Tag	Vr	Type	Comment
Procedure Step State	(0074,1000)	CS		
Procedure Step Progress Information Sequence	(0074,1002)	SQ		
>Procedure Step Progress	(0074,1004)	DS		
>Procedure Step Progress Description	(0074,1006)	ST		

9.1.3 Attribute Mapping

Not applicable

9.1.4 Coerced/Modified Fields

Not applicable

9.2 DATA DICTIONARY OF PRIVATE ATTRIBUTES

All used Private Creators are listed in the table below. Usage of Private Attributes are listed in each module specification.

Attribute name	Tag	VR	VM	Value
ProNova Solutions	(0009,00FF)	LO	1	ProNova Solutions
RaySearch Private Creator	(4001,0010)	LO	1	RAYSEARCHLABS 2.0

9.3 CODE TERMINOLOGY AND TEMPLATES

Not applicable

9.4 GRAYSCALE IMAGE CONSISTENCY

Not applicable

9.5 STANDARD EXTENDED/SPECIALIZED/PRIVATE SOP CLASSES

9.5.1 Standard extended SOP Class

9.5.1.1 RT Ion Plan IOD

Attribute name	Tag	Vr	Type	Comment
Ion Beam Sequence	(300A,03A2)	SQ	1	
>Planned Verification Image Sequence	(300A,00CA)	SQ	3	
>>Imaging Device-Specific Acquisition Parameters	(300A,00CC)	LO	3	

9.5.2 Specialized SOP Class

Not applicable

9.5.3 Private SOP Class

Not applicable

9.6 PRIVATE TRANSFER SYNTAXES

Not applicable



CONTACT INFORMATION



RaySearch Laboratories AB (publ)
Sveavägen 44
SE-111 34 Stockholm
Sweden

Contact details head office

P.O. Box 3297
SE-103 65 Stockholm, Sweden
Phone: +46 8 510 530 00
Fax: +46 8 510 530 30
info@raysearchlabs.com
www.raysearchlabs.com

RaySearch Americas

Phone: +1 877 778 3849

RaySearch France

Phone: +33 975 433 632

RaySearch Korea

Phone: +82 10 2230 2046

RaySearch Belgium

Phone: +32 2 213 83 65

RaySearch Germany

Phone: +49 30 89 36 06 90

RaySearch Singapore

Phone: +65 81 28 59 80

RaySearch China

Phone: +86 137 0111 5932

RaySearch Japan

Phone: +81 3 4405 6902

RaySearch UK

Phone: +44 7508 426 563

